

These vacuum regulators allow the regulation of continuous suction during surgical procedures in tracheal and pharyngeal ways and during continuous gastric drainage.

- The low-vacuum continuous vacuum regulator contains a pressure limiting valve for 140 mmHg.
- Classified as: **Low Vacuum / High Flow**.
- For the patient's security, the regulators are equipped with a positive pressure relief valve, in case of accidental connection to a positive gas pressure.
- It incorporates a **REG-OFF** device that allows the operator to interrupt aspiration without closing the suction control valve.
- Pressing the **PUSH** button, the accessories, like filters or DISS connectors, are released for their easy replacement, ensuring the safety of both the user and the patient (*see PUSH Accessories catalogue*).
- They can be supplied in single or double versions and can be mounted with a direct connector either on a rail fast clamp or on a Ø38,0 mm bar flange. They can be supplied by hose with different European and American connectors (*see Connectors and valves catalogue*).
- The original set includes the regulator, a disposable filter with safety overflow hose, three disposable filters and a metallic connector for flexible hose.



TECHNICAL SPECIFICATIONS

Range:	Low Vacuum: 0-160 mmHg	Flow range: 0-40 l/min
Precision vacuum gauge:	± 3%.	
Dimensions:	147 x 76 x 101 mm	
Weight:	340 g (Without connector).	

	DIN 13260	NF S90-116	BS 5682
Just Vac LV Analog. + Connector	11VPBVAEDN	11VPBVAENF	11VPBVAEBS
Just Vac LV Analog. D/M+ Connector	11VPBAREDN	11VPBARENF	11VPBAREBS
Just Vac LV Analog. to rail - Knob clamp	13VPBVAEPD	13VPBVAEPN	13VPBVAEPB
Just Vac LV Analog. to rail D/M- Knob clamp	13VPBAREPD	13VPBAREPN	13VPBAREPB
Just Vac LV Analog. with Ø38,0 mm tube flange	14VPBVAEPD	14VPBVAEPN	14VPBVAEPB
Just Vac LV Analog. D/M Ø38,0 mm tube flange	14VPBAREPD	14VPBAREPN	14VPBAREPB

¹ For other outputs replace **E** (disposable filter) with: **D** (DISS output), **T** (filter with safety overflow trap), **F** (bottle of 0,5 L), **M** (metal nipple) or **O** (free output).

² For other rail mountings replace **P** (knob clamp) with: **G** (pushbutton clamp), **R** (wall plate) or **H** (bearing plate).

³ Please consult our staff about other gas standards.

⁴ For digital models replace **A** with **D**.



C/ Tungsteno 11-15
 Arganda del Rey, 28500, Madrid
 Telf.: +34 91 871 99 50
 www.itcsal.com itcsal@itcsal.com